

JUN 12 2001

**CD HORIZON® Spinal System**  
**Summary of Safety and Effectiveness**  
**January 2001**

**I. Company:** Medtronic Sofamor Danek, Inc. USA  
1800 Pyramid Place  
Memphis, TN 38132  
(901) 396-3133

**II. Proposed Proprietary Trade Name:** CD HORIZON® Spinal System

**III. Product Description**

The CD HORIZON® Spinal System consists of a variety of rods, hooks, screws, CROSSLINK® plates, staples, and other connecting components used to build a spinal construct. Instrumentation is also available to facilitate implantation of the device components.

Certain implant components from other Medtronic Sofamor Danek spinal systems can be used with the CD HORIZON® Spinal System. These components include TSRH® rods, hooks, screws, plates, CROSSLINK® plates, connectors, staples and washers; GDLH® rods, hooks, connectors and CROSSLINK® bar and connectors; LIBERTY® rods and screws; DYNALOK PLUS® bolts; and Medtronic Sofamor Danek Multi-Axial rods and screws.

**CD HORIZON® hooks are intended for posterior use only. CD HORIZON® staples and CD HORIZON® ECLIPSE® screws are intended for anterior use only. However, for patients of smaller stature, stainless steel CD HORIZON® 4.5mm rods and associated components may be used posteriorly.**

The purpose of this 510(k) submission is to allow surgeons the option of using CD HORIZON® 4.5mm rods and associated components posteriorly.

**IV. Indications**

When used in a percutaneous posterior approach with the SEXTANT instrumentation, the CD HORIZON® Cannulated M8 Multi-Axial Screw components are intended for the following indications:

When used as a posterior spine thoracic/lumbar system, the CD HORIZON® Cannulated M8 Multi-Axial Screw components are intended for: (1) degenerative disc disease (as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), (2) spinal stenosis, (3) spondylolisthesis, (4) spinal deformities (i.e. degenerative scoliosis, kyphosis, and/or lordosis), (5) fracture, (6) pseudarthrosis, (7) tumor resection, and/or (8) failed previous fusion.

In addition, when used as a pedicle screw fixation system the CD HORIZON® Cannulated M8 Multi-Axial Screw components are also indicated for skeletally mature patients: (a) having severe spondylolisthesis (Grades 3 and 4) of the fifth lumbar-first sacral (L5-S1) vertebral joint; (b) who are receiving fusions using autogenous bone graft only; (c) who are having the device fixed or attached to the lumbar and sacral spine (L3 and below); (d) who are having the device removed after the development of a solid fusion mass.

**The CD HORIZON® ECLIPSE® components are intended for the following indications:**

When used as an anterolateral thoracic/lumbar system, the CD HORIZON® Spinal System is intended for the following indications: (1) degenerative disc disease (as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), (2) spinal stenosis, (3) spondylolisthesis, (4) spinal deformities (i.e., scoliosis, kyphosis, and/or lordosis), (5) fracture, (6) pseudarthrosis, (7) tumor resection, and/or (8) failed previous fusion.

The CD HORIZON® system is also intended for the following indications:

When used as a pedicle screw fixation system of the non-cervical posterior spine in skeletally mature patients, the CD HORIZON® Spinal System is indicated for one or more of the following: (1) degenerative spondylolisthesis with objective evidence of neurologic impairment, (2) fracture, (3) dislocation, (4) scoliosis, (5) kyphosis, (6) spinal tumor, and/or (7) failed previous fusion (pseudarthrosis).

In addition, when used as a pedicle screw fixation system, the CD HORIZON® Spinal System is indicated for skeletally mature patients: (a) having severe spondylolisthesis (Grades 3 and 4) of the fifth lumbar-first sacral (L5-S1) vertebral joint; (b) who are receiving fusions using autogenous bone graft only; (3) who are having the device fixed or attached to the lumbar and sacral spine (L3 and below); and (4) who are having the device removed after the development of a solid fusion mass.

When used as a posterior, non-cervical, non-pedicle screw fixation system, the CD HORIZON® Spinal System is intended for the following indications: (1) degenerative disc disease (as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), (2) spinal stenosis, (3) spondylolisthesis, (4) spinal deformities (i.e., scoliosis, kyphosis, and/or lordosis), (5) fracture, (6) pseudarthrosis, (7) tumor resection, and/or (8) failed previous fusion.

**V. Substantial Equivalence**

Documentation was provided which demonstrated the CD HORIZON® Spinal System to be substantially equivalent to itself.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

JUN 12 2001

Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

Richard Treharne, Ph.D.  
Senior Vice President, Regulatory Affairs  
Medtronic Sofamor Danek  
1800 Pyramid Place  
Memphis, Tennessee 38182

Re: K010181/S1  
Trade/Device Name: CD HORIZON® Spinal System  
Regulation Number: 888.3050, 888.3060 and 888.3070  
Regulatory Class: II  
Product Code: KWP, KWQ, MNH and MNI  
Dated: April 18, 2001  
Received: April 19, 2001

Dear Dr. Treharne:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Current Good Manufacturing Practice requirements, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic QS inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4659. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its internet address "<http://www.fda.gov/cdrh/dsma/dsmamain.html>".

Sincerely yours,

A handwritten signature in black ink, appearing to read "Celia M. Witten", is written over a horizontal line.

Celia M. Witten, Ph.D., M.D.  
Director  
Division of General, Restorative  
and Neurological Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

510(k) Number (if known): K010181

Device Name: CD HORIZON® Spinal System

**Indications for Use:**

When used in a percutaneous posterior approach with the SEXTANT instrumentation, the new CD HORIZON® CANNULATED M8 MULTI-AXIAL SCREW components are intended for the following indications:

When used as a posterior spine thoracic/lumbar system, the CD HORIZON® CANNULATED M8 MULTI-AXIAL SCREW components are intended for the following indications: (1) degenerative disc disease (as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), (2) spinal stenosis, (3) spondylolisthesis, (4) spinal deformities (i.e. degenerative scoliosis, kyphosis, and/or lordosis), (5) fracture, (6) pseudarthrosis, (7) tumor resection, and/or (8) failed previous fusion.

In addition, when used as a pedicle screw fixation system, the CD HORIZON® CANNULATED M8 MULTI-AXIAL SCREW components are also indicated for skeletally mature patients: (a) having severe spondylolisthesis (Grades 3 and 4) of the fifth lumbar-first sacral (L5-S1) vertebral joint; (b) who are receiving fusions using autogenous bone graft only; (c) who are having the device fixed or attached to the lumbar and sacral spine (L3 and below); (d) who are having the device removed after the development of a solid fusion mass.

The CD HORIZON® ECLIPSE® components are intended for the following indications:

When used as an anterolateral thoracic/lumbar system, the CD HORIZON® Spinal System is intended for the following indications: (1) degenerative disc disease (as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), (2) spinal stenosis, (3) spondylolisthesis, (4) spinal deformities (i.e., scoliosis, kyphosis, and/or lordosis), (5) fracture, (6) pseudarthrosis, (7) tumor resection, and/or (8) failed previous fusion.

The CD HORIZON® Spinal System is also intended for the following indications:

When used as a pedicle screw fixation system of the non-cervical posterior spine in skeletally mature patients, the CD HORIZON® Spinal System is indicated for one or more of the following: (1) degenerative spondylolisthesis with objective evidence of neurologic impairment, (2) fracture, (3) dislocation, (4) scoliosis, (5) kyphosis, (6) spinal tumor, and/or (7) failed previous fusion (pseudarthrosis).

In addition, when used as a pedicle screw fixation system, the CD HORIZON® Spinal System is indicated for skeletally mature patients: (a) having severe spondylolisthesis (Grades 3 and 4) of the fifth lumbar-first sacral (L5-S1) vertebral joint; (b) who are receiving fusions using autogenous bone graft only; (3) who are having the device fixed or attached to the lumbar and

sacral spine (L3 and below); and (4) who are having the device removed after the development of a solid fusion mass.

When used as a posterior, non-cervical, non-pedicle screw fixation system, the CD HORIZON® Spinal System is intended for the following indications: (1) degenerative disc disease (as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), (2) spinal stenosis, (3) spondylolisthesis, (4) spinal deformities (i.e., scoliosis, kyphosis, and/or lordosis), (5) fracture, (6) pseudarthrosis, (7) tumor resection, and/or (8) failed previous fusion.

(PLEASE DO NOT WRITE BELOW THIS LINE—CONTINUE ON ANOTHER PAGE IF NEEDED)

\_\_\_\_\_  
Concurrence of CDRH, Office of Evaluation (ODE)

Prescription Use X  
(Per 21 CFR 801.109)

OR

Over-the-counter Use \_\_\_\_\_

(Optional 1-2-96)

*Thomas H. ...*  
(Division Sign-Off)  
Division of General, Restorative  
and Neurological Devices

510(k) Number K010181